### Acceptance Protocol

**NMR** Spectrometer



#### **Customer Information**

Customer NameProf. Steven BrownOperator NameDr Trent FranksCompanyUniversity of Warwick

**Address** 

Postal Code / City / Country Coventry, UK

**Phone Contact Customer** +44 (0)7512855361

Fax

**E-Mail** t.franks@warwick.ac.uk

**Bruker Information** 

OfficeCoventryEngineerAriana JonesCentral Hotline Phone0247 6855333

Central Hotline E-Mail service.bbio.uk@bruker.com

**Spectrometer Information** 

**Order No.** 408457

Contract No.

System Avance Neo 1000

Console Part and Serial No HCAB-20 / 5

Coil
Dewar
Location

**TopSpin Version** TopSpin 4.0.9 - Build 597

## **Acceptance**

I, an authorized customer representative, acknowledge that the above referenced system was installed and demonstrated to operate in accordance with the specifications mutually agreed upon by both parties. We accept the delivery and installation of this system as complete (except for items excluded below) and release Bruker from any further obligation, other than those obligations as specified during the warranty period. With this signature, the warranty period for non-excluded items commences according to the contractual agreement.

The warranty starts on Dec 22, 2020.

Place: Coventry, UK Place: Coventry, UK

Date: Dec 22, 2020 Date: Dec 22, 2020

Prof. Steven Brown Ariana Jones

Customer Representative Signature Bruker Representative Signature

ZFPT0008 Bruker BioSpin Report Name: 0.7mm\_Install Index: 02 Page: 1 / 4

#### **NMR Probe**

DescriptionProbe IDInspection LotStatusPH MASDVT1000S6 BL0.7 N/D/C/H NO\_I/EH171374\_00010.7mm\_Installpass

Copies of all spectra (default and additional) are included in customer's PDF report.

### **Installation Checklist**

mstanation oncernst	
Installation All connections and grounding All firmware Cortab for required nuclei Lift / spin calibration Software licenses He / N2 log files activated MICS installed	pass fail n/a
Customer Training Basic safety Magnet safety and refilling Handling of cryogenic liquids Hardware overview Console on/off operation Basic operation Troubleshooting Backup (nmr_save, Images) Introduction to IconNMR Assure-SST / Performance check CryoProbe Handling / cleaning of probe He cylinder exchange He compressor cooling RF heating / power limits RF routing	pass fail n/a
Acceptance and Warranty Explaination of warranty Spectrometer documentation Customer support hotlines	pass fail n/a  ② □ □  ② □ □

Optional Components	pass	fail	n/a
Sample Changer		Π.	,∝
MAS controller	$\overline{}$	$\overline{\Box}$	Ø
		$\exists$	Ø
High power equipment			
LC-NMR			$\checkmark$
Liquid Handler SamplePro Tube			$\checkmark$
Micro-Imaging			$\checkmark$
Diffusion			$\checkmark$
CryoProbe / Cryoplatform			$\checkmark$
BNL / BSNL			$\checkmark$
Additional cooling/heating units (like			$\checkmark$
BCU1 / BCU2)			
LT-MAS (Low Temperature MAS			$\checkmark$
equipment)			
Gyrotron magnet and DNP console			$\checkmark$

ZFPT0008 Bruker BioSpin Report Name: 0.7mm\_Install Index: 02 Page: 2 / 4

### **General Test Information**

### **Summary of Inspection Lot**

Description **Probe ID** Inspection Lot PH MASDVT1000S6 BL0.7 N/D/C/H NO\_I/E H171374\_0001 0.7mm\_Install

### **Experiments Measured**

Sample	Experiment	Status
Z163271	Magic Angle setting, MAS (NPT_79Br_MAS_magicAngle)	pass
Z163271	Maximum spin rate testing, MAS (NPT_79Br_MAS_maxSpinRate)	pass
Z163271	Optimization of 79Br frequency (NPT_79Br_MAS_fieldsetting)	pass
Z163274	Optimization of 13C frequency (NPT_13C_MAS_fieldsetting_dec1h)	pass
Z163274	P90 1H pulse calibration, MAS (NPT_1H_MAS_p90det_1h)	pass
Z163271	P90 79Br pulse calibration, MAS (NPT_79Br_MAS_p90det_79br)	pass
Z163276	P90 13C 1H-13C CP pulse calibration, MAS (NPT_13C_MAS_p90det_cp1h_13c)	pass
Z163274	P90 13C pulse calibration, MAS (NPT_13C_MAS_p90det_13c)	pass
Z163276	P90 15N 1H-15N CP pulse calibration, MAS (NPT_15N_MAS_p90det_cp1h_15n)	pass
Z163275	CP 1H-13C sensitivity, MAS (NPT_13C_MAS_sino_cp1h_13c)	pass
Z163274	13C sensitivity, MAS (NPT_13C_MAS_sino_13c)	pass
Z163274	1H sensitivity, MAS (NPT_1H_MAS_sino_1h)	pass
Z163276	Double CP 1H-15N-13C, MAS (NPT_13C_MAS_double_cp1h15n_13c)	pass
Z163276	CP 1H-13C parameter optimization, MAS (NPT_13C_MAS_paropt_cp1h_13c)	pass
Z163276	CP 1H-15N parameter optimization, MAS (NPT_15N_MAS_paropt_cp1h_15n)	pass

## **Achieved Specifications**

#### **Pulse Width**

Nucleus	Sample		90° P	ulse	Power Limit	Mathad	Status	
Nucleus	Sample		Duration [µs]	Power [W]	[W]	Method	Status	
<sup>1</sup> H	Z163274	spec.	1.50	-	25	direct	2000	
'П	2103274	ach.	1.44	15.2	25	allect	pass	
13C	7462074	spec.	3.00	-	20	dino ot	2000	
130	Z163274	ach.	2.84	19.1	30	direct	pass	
13C	Z163276	spec.	3.50	-	30	with CP	pass	
	2103276	ach.	3.40	12.6	30			
15 <b>N</b>	7462076	spec.	4.00	-	60	with CP	2000	
1914	Z163276	ach.	3.95	51.2	60		pass	
79 <b>B</b> r	7162071	spec.	3.50	-	20	direct	2000	
, .DI	Z163271	ach.	3.37	14.0	30	direct	pass	

#### Sensitivity

Nucleus	Sample		S/N Remarks		Status
13C	7462075	spec.	-	sensitivity of <sup>1</sup> H- <sup>13</sup> C cross-polarization	2000
150	Z163275	ach.	7.0	sensitivity of 'H-19C cross-polarization	pass

#### Sensitivity with NS

Nucleus	Sample		S/N	FWHM [Hz]	NS	Remarks	Status
		spec.	-	-	-	noise: 20 ppm	
<sup>1</sup> H	Z163274	ach.	4812.8	93.1	1	variable, method: sino best	pass
		spec.	-	7.0	-	noise: 20 ppm	
13 <b>C</b>	Z163274	ach.	8.2	3.5	1	variable, method: sino best	pass

# Samples used for Inspection Lot

Sample	Description			
Z163271	otassium Bromide (KBr, 0.5 ul)			
Z163274	Adamantane (0.5 ul)			
Z163275	Alpha-glycine (0.5 ul)			
Z163276	2-13C, 15N alpha-glycine (0.5 ul)			

# **Remarks / Exclusions**

ZFPT0008 Bruker BioSpin Index: 02