

Risk Assessment Summary Report/Print (landscape)



Reference	3502	Description of Space or Activity/Task or Equipment	Use of the OmniScan Profilm 2D white light interferometer located in millburn house G31
Assessment Date	06/09/2024	Publish To Portal	No
Assessor Name	Ben Breeze	Risk Assessment Title	OmniScan Profilm 2D white light interferometer
Assessment Team Members		Review Date	No Review Set
Role / Space / Project Reference		Current Risk Level (1=Very Low, 2=Low, 3=Moderate, 4=High, 5=Very High)	2
Department	Use the search function above or double click here for org chart -> Academic Faculties -> Faculty of Science, Engineering and Medicine -> Research Technology Platforms - RTPs -> Spectroscopy Research Technology Platform - RTP	Final Risk Level (1=Very Low, 2=Low, 3=Moderate, 4=High, 5=Very High)	2
Location Details	01.005.000.031 Millburn House Ground Floor G31 Optical Microscope Room	Risk Assessment Number	0
Risk Assessment Category	Equipment	Additional Information	A separate Risk assessment must be carried out by the user for each sample being used these must be logged on the online form prior to experiments https://warwick.ac.uk/research/rtp/spectroscopy/Safety/Sample These individual sample assessments are available from B Breeze or B Green on request. The Diamond view should not be operated before training.
Date Record Created	06/09/2024		

Hazard Type & Hazard Description	Who may be at Risk? & How May Person(s) Be Harmed	Existing Control Measures	L	S	R	Where current risk is M, H or VH, what additional Control Measures are required?	L	S	R
Equipment/Plant Contact with rotating mechanical parts (Friction). Contact with mechanical pinch or nip points.	Laboratory Worker Staff Student Cuts bruises	Admin - Safe System of Work (SSoW) established and trained out. Admin - Specific training provided to equipment operators. Eng- Loos hair to be tied back Eng - Loose clothing avoided	Minor	Unlikely	Very Low		Minor	Unlikely	Very Low

Risk Assessment Summary Report/Print (landscape)



Equipment/Plant Broken glassware	Laboratory Worker Staff Student cuts	Admin - Glassware to be disposed of in the correct waste stream Admin - Specific training provided for uses PPE - Personal Protective Equipment (PPE) is issued and worn (specify).	Minor	Possible	Low		Minor	Possible	Low
Electricity Contact with live electrics.	Laboratory Worker Staff Student Electric Shock / Burns	Electric Burns Shocks	Major	Unlikely	Low		Major	Unlikely	Low
Assessment Conclusion	Very Safe Instrument								